Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category	Bioanalytical Standard
Product Model Number	2778
Lot Number	19J100340
Method	Comparative enzymatic testing of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
SUCROSE	(Average of production) - (average of metrology)	± 0.3 g/L	0.18 g/L
	Coefficient of variation for production standard (CV, %)	≤ 2 %	1.19%
	(Average of production) - (average of metrology)	N/A	N/A
	Coefficient of variation for production standard (CV, %)	N/A	N/A

Conforms to acceptance criteria

David Stoler

Completed by:

Date:

09-17-2019



mullel

Lisa M. Abel Director, Quality and Corporate Responsibility