## **Certificate of Analysis**



**YSI Incorporated** 

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category	Bioanalytical Standard	
Product Model Number	2790	<u>_</u>
Lot Number	19J100507	
Method	Comparative enzymatic test	ng of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
ETHANOL STD A 2.0G/L	(Average of production) - (average of metrology)	±40 mg/L	7.05 mg/L
	Coefficient of variation for production standard (CV, %)	≤2%	1.2%
ETHANOL STD B 3.2G/L	(Average of production) - (average of metrology)	±64 mg/L	17.5 mg/L
	Coefficient of variation for production standard (CV, %)	≤2%	0.9%

Conforms to accep	otance criteria		
Completed by:	David Stoler	Date:	9-26-2019



Lisa M. Abel Director, Quality and Corporate Responsibility

Tarralel