## **Certificate of Analysis**



**YSI Incorporated** 

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category	Bioanalytical Standard	
Product Model Number	2726	<u>_</u>
Lot Number	19J100709	<u> </u>
Method	Comparative enzymatic testi	ng of Production standard vs. Metrology standard

<b>Chemistry Tested</b>	Determination	Tolerance	Result
Methanol Std A	(Average of production) - (average of	± 20 mg/L	6.6 mg/L
1.0 g/L	metrology)		
	Coefficient of variation for production	≤ 2 %	1.1%
	standard (CV, %)		
Methanol Std B	(Average of production) - (average of	$\pm$ 50 mg/L	1.1 mg/L
2.50 g/L	metrology)		
	Coefficient of variation for production	≤ 2 %	0.7%
	standard (CV, %)		

Conforms to acceptance criteria				
Completed by:	Lucas Stover	Date:	10/02/2019	



Lisa M. Abel Director, Quality and Corporate Responsibility

Thralel