Certificate of Analysis



Product Category

Lot Number

Product Model Number

YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

2784

19K100079

Bioanalytical Standard

standard (CV, %)

ethod	Comparative enzymatic testing of Production standard vs. Metrology standard			
Chemistry Tested	Determination	Tolerance	Result	
LACTOSE	(Average of production) - (average of metrology)	±0.3 g/L	0.0 g/L	
	Coefficient of variation for production standard (CV, %)	≤2%	1.3%	
	(Average of production) - (average of metrology)	N/A	N/A	
	Coefficient of variation for production	N/A	N/A	

Conforms to acceptance criteria					
Completed by:	Lucas Stover		Date:	10/11/2019	



Lisa M. Abel Director, Quality and Corporate

Turalel

Responsibility