Certificate of Analysis



Product Category

Lot Number

Product Model Number

YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

2778

19K100444

Bioanalytical Standard

standard (CV, %)

thod	Comparative enzymatic testing of Production standard vs. Metrology standard			
Chemistry Tested	Determination	Tolerance	Result	
SUCROSE	(Average of production) - (average of metrology)	± 0.3 g/L	0.05 g/L	
	Coefficient of variation for production standard (CV, %)	≤ 2 %	0.5%	
	(Average of production) - (average of metrology)	N/A	N/A	
	Coefficient of variation for production	N/A	N/A	

Conforms to acceptance criteria					
Completed by:	Lucas Stover		Date:	10/21/2019	



Lisa M. Abel Director, Quality and Corporate Responsibility

Turalel