## **Certificate of Analysis**



**YSI Incorporated** 

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category	Bioanalytical Standard	
Product Model Number	7141	
Lot Number	19K100445	
Method	Comparative enzymatic testing of Production standard vs. Metrology standard	

<b>Chemistry Tested</b>	Determination	Tolerance	Result
Glycerol	(Average of production) - (average of metrology)	± 0.30 g/L	0.00g/L
	Coefficient of variation for production standard (CV, %)	≤ 2 %	1.1 %
None	(Average of production) - (average of metrology)	N/A	N/A
	Coefficient of variation for production standard (CV, %)	N/A	N/A

Conforms to acceptance criteria							
Completed by:	Lucas Stover	Date:	10/21/2019				



Lisa M. Abel Director, Quality and Corporate Responsibility

Thralel