Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Method	Comparative enzymatic testi	ng of Production standard vs. Metrology standard
Lot Number	19L100256	_
Product Model Number	7141	<u> </u>
Product Category	Bioanalytical Standard	

Chemistry Tested	Determination	Tolerance	Result
Glycerol	(Average of production) - (average of metrology)	± 0.30 g/L	0.02g/L
	Coefficient of variation for production standard (CV, %)	≤ 2 %	0.51 %
None	(Average of production) - (average of metrology)	N/A	N/A
	Coefficient of variation for production standard (CV, %)	N/A	N/A

Conforms to acce	ptance criteria		
Completed by:	David Stoler	Date:	11-20-2019



Lisa M. Abel Director, Quality and Corporate Responsibility

Turalel