Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category	Bioanalytical Standard	
Product Model Number	7142	
Lot Number	19L100257	
Method	Comparative enzymatic testing of Production	n standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
Glycerol	(Average of production) - (average of metrology)	± 0.48 g/L	0.02 g/L
	Coefficient of variation for production standard (CV, %)	$\leq 2 \%$	0.51%
	(Average of production) - (average of metrology)	N/A	N/A
	Coefficient of variation for production standard (CV, %)	N/A	N/A

Conforms to acceptance criteria

David Stoler

Completed by:

Date:

11-18-2019



mullel

Lisa M. Abel Director, Quality and Corporate Responsibility