## **Certificate of Analysis**



## **YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category	Bioanalytical Standard	
Product Model Number	2790	
Lot Number	19L100504	
Method	Comparative enzymatic testing	of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
ETHANOL STD A 2.0G/L	(Average of production) - (average of metrology)	±40 mg/L	0 mg/L
	Coefficient of variation for production standard (CV, %)	≤2%	0.4%
ETHANOL STD B 3.2G/L	(Average of production) - (average of metrology)	±64 mg/L	2 mg/L
	Coefficient of variation for production standard (CV, %)	≤2%	1.4%

Conforms to acceptance criteria

Completed by:

Lucas Stover

Date:

11/26/2019



mullel

Lisa M. Abel Director, Quality and Corporate Responsibility