Certificate of Analysis



YSI Incorporated Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category	ISE Buffer			
Product Model Number	2970			
Lot Number	19M100218			
Method	Comparative ISE testing of Pr	oduction buffer vs. Met	trology buffer	

Chemistry Tested	Determination	Tolerance	Result
AMMONIUM	(Average of production) - (average of metrology)	<.250 mV	0.010 mV
	Coefficient of variation for production buffer (CV, %)	≤ 2.0%	0.51%
POTASSIUM	(Average of production) – (average of metrology)	< .250 mV	0.074 mV
	Coefficient of variation for production buffer (CV, %)	≤ 2.0%	0.81%

Conforms to acceptance criteria							
Completed by:	David Stoler	Date:	12-12-2019				

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp Quality Manager, Analytics

Dregory W. Popp