## **Certificate** of Analysis



## **YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category	ISE Standard	
Product Model Number	2972	
Lot Number	19M100219	
Method	Comparative ISE testing of Pr	oduction standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
Ammonium	(Average of production) - (average of metrology)	< .250 mV	.069 mV
	Coefficient of variation for production standard (CV, %)	≤ 2.0 %	0.12 %
Potassium	(Average of production) - (average of metrology)	< .250 mV	.124 mV
	Coefficient of variation for production standard (CV, %)	≤ 2.0 %	0.38 %

Conforms to acceptance criteria

Completed by: David Stoler

Date: 12-17-2019

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Dregory W. Popp

Gregory W. Popp Quality Manager, Analytics