Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category	Bioanalytical Standard				
Product Model Number	2368				
Lot Number	19M100444				
Method	Comparative enzymatic testin	g of Production	standard vs.	Metrology	standard

Chemistry Tested	Determination	Tolerance	Result
Glucose	(Average of production) - (average of metrology)	± .13 mmol/L	.05 mmol/L
	Coefficient of variation for production standard (CV, %)	≤ 2.0%	0.27%

Conforms to acce	otance criteria		
Completed by:	David Stoler	Date:	1-8-2020

 $The \ procedures \ used to \ manufacture \ and \ test \ this \ product \ have \ been \ verified \ by \ QA for \ GMP \ compliance.$

STOT CERTIFICATION OF THE PROPERTY OF THE PROP

Gregory W. Popp Quality Manager, Analytics

Dregory W. Popp