Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category Buffer and solid products

Product Model Number 2357

Lot Number 200109GFS098005

Method Verification of mass and pH (for buffers)

| Determination | Specification | Result |
|---------------------------|----------------|--------|
| Average mass | 6.00 - 6.70 g | 6.32g |
| Average pH (buffers only) | 7.1-7.5 | 7.2 |

Conforms to acceptance criteria

Completed by: Lucas Stover Date: 02/05/2020



Lisa M. Abel Director, Quality and Corporate Responsibility

Turalel