Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category	Buffer and solid products

Product Model Number 2357

Lot Number 200203GFS098005

Method Verification of mass and pH (for buffers)

Determination	Specification	Result
Average mass	$6.35g \pm 0.35g$	6.34g
Average pH (buffers only)	7.3 ± 0.2	7.2

Conforms to accepta	ance criteria			
Completed by:	David Stoler	Date:	3-11-2020	

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp Quality Manager, Analytics

Dregory W. Popp