Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category	Bioanalytical Standard				
Product Model Number	2801				
Lot Number	20A100253				
Method	Comparative enzymatic testin	g of Production	standard vs.	Metrology	standard

Chemistry Tested	Determination	Tolerance	Result
Glucose	(Average of production) - (average of metrology)	± 5 mg/L	0.15 mg/L
	Coefficient of variation for production standard (CV, %)	≤ 2.0 %	0.82 %

Conforms to acce	ptance criteria		
Completed by:	David Stoler	Date:	1-15-20

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.

O CERTIFICATION OF THE PROPERTY OF THE PROPERT

Gregory W. Popp Quality Manager, Analytics

Dregory W. Popp