Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category	Bioanalytical Standard
Product Model Number	2367
Lot Number	20A100863
Method	Comparative enzymatic testing of Produc

Comparative enzymatic testing of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
GLUCOSE	(Average of production) - (average of	$\pm 0.05 \text{ mmol/L}$	0.03 mmol/L
	metrology)		
	Coefficient of variation for production	$\leq 2\%$	0.59%
	standard (CV, %)		

Conforms to acceptance criteria					
Completed by:	David Stoler		Date:	1-28-2020	



Dregory W. Popp

Gregory Popp Quality Manager