## **Certificate of Analysis**



**YSI Incorporated** 

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Method	Comparative enzymatic testing of Production standard vs. Metrology standard		
Lot Number	20A100866	_	
Product Model Number	2768	_	
Product Category	Bioanalytical Standard		

<b>Chemistry Tested</b>	Determination	Tolerance	Result
Xylose	(Average of production) - (average of metrology)	± 1.5 g/L	0.2 g/L
	Coefficient of variation for production standard (CV, %)	≤ 2 %	0.1.1 %
None	(Average of production) - (average of metrology)	N/A	N/A
	Coefficient of variation for production standard (CV, %)	N/A	N/A

Conforms to acceptance criteria						
Completed by:	Lucas Stover	Date:	02/04/2020			



Lisa M. Abel Director, Quality and Corporate

Thralel

Responsibility