## **Certificate of Analysis**



**YSI Incorporated** 

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category Membrane

Product Model Number 2903

Lot Number 20A100889

Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	5-40 nA	26.7 nA
	Average Exclusion	0.00±0.5g/L	0.07 g/L

Conforms to acceptance criteria

Completed by: Christy Woodford 3/10/2020

Your Name



Gregory W. Popp Quality Manager Analytics

Dregory W. Popp