## **Certificate of Analysis**



## **YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category	Bioanalytical Standard	
Product Model Number	2748	
Lot Number	20B130008	
Method	Comparative enzymatic testing	of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
Dextrose	(Average of production) - (average of metrology)	± 0.2 g/L	0.04 g/L
	Coefficient of variation for production standard (CV, %)	≤ 2.0%	0.72%
Lactate	(Average of production) - (average of metrology)	± 0.02 g/L	0.02 g/L
	Coefficient of variation for production standard (CV, %)	≤ 2.0%	0.66%

Conforms to acceptance criteria

Completed by: David Stoler

Date: 2-17-2020

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Dregory W. Popp

Gregory W. Popp Quality Manager, Analytics