Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category Membrane

Product Model Number 2329

Lot Number 20B130015

Method Amperometric testing with standards

| Time of Testing | Determination | Specification | Results |
|-----------------|---------------------------------|------------------|--------------|
| Initial | Average Calibration Sensitivity | 6.0-30.0 nA | 13.1 nA |
| | Average Linearity | 14.3-15.7mmol/L | 14.92 mmol/L |
| | Potassium Ferrocyanide | -0.3 - 0.3mmol/L | 0.00 mmol/L |

| Conforms to acceptance criteria | | |
|---------------------------------|--|--|
| | | |
| | | |

Christy Woodford

Completed by:
Your Name

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Lisa M. Abel Director, Quality and Corporate Responsibility

2/18/2020