Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category Membrane

Product Model Number 2329

Lot Number 20J100306

Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	6.0-30.0 nA	12.2 nA
	Average Linearity	14.3-15.7mmol/L	14.97 mmol/L
	Potassium Ferrocyanide	-0.3 - 0.3mmol/L	0.00 mmol/L

Conforms	to	acceptance	criteria
COLLIGITIS	w	acceptance	Unitoria

Completed by: Christy Woodford 9/14/2020

Your Name

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.

ODON CERTIFICATION OF THE PROPERTY OF THE PROP

Gregory W. Popp Quality Manager Analytics

Dregory W. Popp