Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category	Bioanalytical Standard				
Product Model Number	2748				
Lot Number	20K100079				
Method	Comparative enzymatic testin	g of Production	standard vs. M	letrology	standard

Chemistry Tested	Determination	Tolerance	Result
Lactate	(Average of production) - (average of	± 0.02 g/L	0.01 g/L
	metrology)		
	Coefficient of variation for production standard (CV, %)	≤ 2.0%	0.85%
Glucose	(Average of production) - (average of metrology)	± 0.2 g/L	0.07 g/L
	Coefficient of variation for production standard (CV, %)	≤ 2.0%	0.80%

Conforms to acc	eptance criteria				_
Completed by:	Nick Rittenhouse	Date	:	10/07/2020	
	Amanda Fenner		·		

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp Quality Manager, Analytics

Dregory W. Popp