Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category	Buffer and solid products	
Product Model Number	1579	
Lot Number	20K100266	<u>.</u>
Method	Verification of mass and pH ((for buffers)

Determination	Specification	Result
Average mass	N/A	
Average pH (buffers only)	9.00 ± 0.10	8.99

Conforms to acceptance criteria

Completed by:

Nick Rittenhouse

Date:

10/15/2020

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Dregory W. Popp

Gregory W. Popp Quality Manager, Analytics