## **Certificate of Analysis**



**YSI Incorporated** 

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category	Bioanalytical Standard		
Product Model Number	2790	<u>_</u>	
Lot Number	20K100717	<u> </u>	
Method	Comparative enzymatic test	ing of Production standard vs. M	Metrology standard

Chemistry Tested	De te rmination	Tolerance	Result
Ethanol, 2.0 g/L	(Average of production) - (average of	± 40 mg/L	13.25 mg/L
	metrology)		
	Coefficient of variation for production	≤ 2.0%	1.18%
	standard (CV, %)		
Ethanol, 3.2 g/L	(Average of production) - (average of	± 64 mg/L	20.00 mg/L
	metrology)		
	Coefficient of variation for production	≤ 2.0%	0.64%
	standard (CV, %)		

Conforms to acceptance criteria					
Completed by:	Nick Rittenhouse	Date:	10/27/2020		
	Amanda Fenner				

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp Quality Manager, Analytics

Dregory W. Popp