## **Certificate of Analysis**



**YSI Incorporated** 

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

| Product Category     | Bioanalytical Standard     |   |     |
|----------------------|----------------------------|---|-----|
| Product Model Number | 2790                       | <u> </u>  |     |
| Lot Number           | 20L100414                  | <u> </u>  |     |
| Method               | Comparative enzymatic test | ing of Production standard vs. Metrology standa | .rd |

| Chemistry Tested | Determination  | Tolerance | Result    |
|------------------|--|-----------|-----------|
| Ethanol, 2.0 g/L | (Average of production) - (average of metrology)         | ± 40 mg/L | 6.20 mg/L |
|                  | Coefficient of variation for production standard (CV, %) | ≤ 2.0%    | 0.41%     |
| Ethanol, 3.2 g/L | (Average of production) - (average of metrology)         | ± 64 mg/L | 20.3 mg/L |
|                  | Coefficient of variation for production standard (CV, %) | ≤ 2.0%    | 0.65%     |

| Conforms to acceptance criteria |                  |       |            |  |  |
|---------------------------------|------------------|-------|------------|--|--|
| Completed by:                   | Nick Rittenhouse | Date: | 11/18/2020 |  |  |
|                                 | Amanda Fenner    |       |            |  |  |

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp Quality Manager, Analytics

Dregory W. Popp