Certificate of Analysis



YSI Incorporated Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category	Buffer and solid products
Product Model Number	1579
Lot Number	20L100649
Method	Verification of mass and pH (fo

Determination	Specification	Result
Average mass	N/A	
Average pH (buffers only)	9.00 ± 0.10	8.99

Conforms to acce	eptance criteria			
Completed by:	Nick Rittenhouse	Date:	11/24/2020	
	Amanda Fenner			

 $The \ procedures \ used to \ manufacture \ and \ test \ this \ product \ have \ been \ verified \ by \ QA \ for \ GMP \ compliance.$



Gregory W. Popp Quality Manager, Analytics

Dregory W. Popp