Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category	Buffer and solid products		
Product Model Number	1579	-	
Lot Number	20M100169	-	
Method	Verification of mass and pH ((for buffers)	

Determination	Specification	Result
Average mass	N/A	
Average pH (buffers only)	9.00 ± 0.10	8.97

Conforms to acceptance criteria						
Completed by:	Amanda Fenner	Date:	12/11/2020			

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Dregory W. Popp

Gregory W. Popp Quality Manager, Analytics