Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Buffer and solid products	
1579	
21K100447	
Verification of mass and pH (for buffers)
	1579 21K100447

Determination	Specification	Result
Average mass	N/A	
Average pH (buffers only)	9.00 ± 0.10	8.94

Conforms to acceptance criteria				
Completed by:	Nick Rittenhouse	Date:	10/21/2021	

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Dregory W. Popp

Gregory W. Popp Quality Manager, Analytics