Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category	Bioanalytical Standard		
Product Model Number	2790	-	
Lot Number	21K100454	_	
Method	Comparative enzymatic testing	ng of Production standard vs. Metrolog	gy standard

Chemistry Tested	Determination	Tolerance	Result
Ethanol, 2.0 g/L	(Average of production) - (average of metrology)	± 40 mg/L	9.60 mg/L
	Coefficient of variation for production standard (CV, %)	≤ 2.0%	0.55%
Ethanol, 3.2 g/L	(Average of production) - (average of metrology)	± 64 mg/L	3.80 mg/L
	Coefficient of variation for production standard (CV, %)	≤ 2.0%	0.75%

Conforms to acce	eptance criteria			
Completed by:	Amanda Fenner	Date:	10/22/2021	

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp Quality Manager, Analytics

Dregory W. Popp