Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category Buffer and solid products

Product Model Number 1579

Lot Number 21L100460

Method Verification of mass and pH (for buffers)

Determination	Specification	Result
Average mass	N/A	
Average pH (buffers only)	9.00 ± 0.10	8.95

Conforms to acceptance criteria					
Completed by:	Amanda Fenner	Date:	11/24/2021		

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp Quality Manager, Analytics

Dregory W. Popp