Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product CategoryMembraneProduct Model Number2329Lot Number21M100164MethodAmperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	6.0-30.0 nA	12.0 nA
	Average Linearity	14.3-15.7mmol/L	15.08 mmol/L
	Potassium Ferrocyanide	-0.3 - 0.3mmol/L	0.00 mmol/L

Conforms to acceptance criteria

Completed by:	Sandy Crawford	12/6/2021
	Your Name	

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Dregory W. Popp

Gregory W. Popp Quality Manager Analytics