Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category	Membrane
Product Model Number	2702
Lot Number	22F100495
Method	Amperometric testing with standards

Time of Testing	Determination	Specification	Result
Initial	Average calibration sensitivity	10-30 nA	19.22 nA
	Average Linearity	23.75-26.25g/L	24.6 g/L
	Potassium Ferrocyanide interference	N/A	N/A
Day =	Average calibration sensitivity	N/A	N/A
	Average Linearity		N/A
	Potassium Ferrocyanide interference	N/A	N/A

Conforms to acceptance criteria

Completed by:

Sandy Crawford

Date:

6/27/22

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Dregory W. Popp

Gregory W. Popp Quality Manager, Analytics