Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category	Bioanalytical Standard	
Product Model Number	2790	
Lot Number	22G100036	
Method	Comparative enzymatic testin	g of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
Ethanol, 2.0 g/L	(Average of production) - (average of metrology)	\pm 40 mg/L	14.65 mg/L
	Coefficient of variation for production standard (CV, %)	≤ 2.0%	0.78%
Ethanol, 3.2 g/L	(Average of production) - (average of metrology)	\pm 64 mg/L	15.40 mg/L
	Coefficient of variation for production standard (CV, %)	≤ 2.0%	0.40%

Conforms to acceptance criteria

Completed by: Amanda Fenner

Date: 7/08/2022

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Dregory W. Popp

Gregory W. Popp Quality Manager, Analytics