## **Certificate of Analysis**



## **YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category	Check Solution
Product Model Number	7173
Lot Number	22K100394
Method	Comparative conductivity testing of Production standard vs. Metrology standard

<b>Chemistry Tested</b>	Determination	Tolerance	Result
Sodium	(Average of production) - (average of metrology)	± 1.0%	0.79%
	Coefficient of variation for production standard (CV, %)	≤ 2.0%	0.48%

Conforms to acceptance criteria						
Completed by:	Amanda Fenner	Date:	10/21/2022			

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp Quality Manager, Analytics

Dregory W. Popp