## **Certificate of Analysis**



## **YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

| Product Category     | Bioanalytical Standard |  |
|----------------------|------------------------|--|
| Product Model Number | 2790                   |  |

Lot Number 22L100233

Method Comparative enzymatic testing of Production standard vs. Metrology standard

| <b>Chemistry Tested</b> | Determination  | Tolerance | Result     |
|-------------------------|--|-----------|------------|
| Ethanol, 2.0 g/L        | (Average of production) - (average of metrology)         | ± 40 mg/L | 2.20 mg/L  |
|                         | Coefficient of variation for production standard (CV, %) | ≤ 2.0%    | 0.78%      |
| Ethanol, 3.2 g/L        | (Average of production) - (average of metrology)         | ± 64 mg/L | 27.55 mg/L |
|                         | Coefficient of variation for production standard (CV, %) | ≤ 2.0%    | 0.46%      |

| Conforms to accep | tance criteria |       |            |
|-------------------|----------------|-------|------------|
| Completed by:     | Amanda Fenner  | Date: | 11/22/2022 |

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp Quality Manager, Analytics