Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category	Buffer and solid products	
Product Model Number	2357	
Lot Number	230719GFS098005	
Method	Verification of mass and pH (for buffers)	

Determination	Specification	Result
Average mass	6.00 – 6.70 g	6.22 g
Average pH (buffers only)	7.1-7.5	7.3

Conforms to acceptance criteria						
Completed by:	Amanda Fenner		Date:	7/26/2023		

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Dregory W. Popp

Gregory W. Popp Quality Manager, Analytics