Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category Buffer and solid products

Product Model Number 2705

Lot Number 230927GFS002699

Method Verification of mass and pH (for buffers)

Determination	Specification	Result
Average mass	6.35 ± 0.35 g	6.18 g
Average pH (buffers only)	7.3 ± 0.2	7.3

Conforms to accepta	ance criteria			
Completed by:	Nick Rittenhouse	Date:	10/2/2023	

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp Quality Manager, Analytics

Dregory W. Popp