Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category	Buffer and solid products		
Product Model Number	2357		
Lot Number	230929GFS098005		
Method	Verification of mass and pH (for buffers		

Determination	Specification	Result
Average mass	6.00 – 6.70 g	6.23 g
Average pH (buffers only)	7.1-7.5	7.3

Conforms to acceptance criteria					
Completed by:	Amanda Fenner	Date	e: <u>10/04/2023</u>		

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Dregory W. Popp

Gregory W. Popp Quality Manager, Analytics