## **Certificate of Analysis**



YSI Incorporated Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category Buffer and solid products

Product Model Number 2357

Lot Number 231006GFS098005

Method Verification of mass and pH (for buffers)

Determination	Specification	Result
Average mass	6.00 - 6.70  g	6.23 g
Average pH (buffers only)	7.1-7.5	7.4

Conforms to acceptance criteria					
Completed by:	Nick Rittenhouse	Date:	10/25/2023		

 $The \ procedures \ used to \ manufacture \ and \ test \ this \ product \ have \ been \ verified \ by \ QA \ for \ GMP \ compliance.$ 



Gregory W. Popp Quality Manager, Analytics

Dregory W. Popp