Certificate of Traceability

YSI Incorporated
1700/1725 Brannum Lane
Yellow Springs, OH 45387 USA
Phone: (937) 767-7241 Toll-Free: (800) 765-4974
Fax: (937) 767-9353

a xylem brand

YSI Product Information

Product Number	098092
Product Description	YSI 2356 D-Glucose Standard 500 mg/dL
Lot Number	26K100273
Product Expiration Date	**10/16/27**

Metrology Primary Standard Information

(Related to Preparation of Product Above)

YSI Metrology Primary Standard ID Number	632356
Concentration of Metrology Standard	D-Glucose, 5 g/L [500 mg/dL]
Date of Preparation	**10/12/26**
Metrology Standard Expiration Date	**4/12/27**

YSI Metrology Materials & Equipment

	Certification ID #	Certificate on File	Date of Last Calibration
Glucose	**917d**	Yes	**1/18/23**
Volumetric Flask (Class A)	**2560**	Yes	**1/30/19**
Volumetric Flask (Class A)	**1843**	Yes	**8/23/08**
Analytical Balance	**6952**	Yes	**7/13/26**
Temperature Probe	**7559**	Yes	**4/21/26**

Conforms to all documented processes for the preparation of YSI Metrology Primary Standards.

Completed By: Amanda Fenner Digitally signed by Amanda Fenner Date: 2026.10.21 11:31:46 -04'00' Date: **10/21/2026**