Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category Membrane

Product Model Number 2329

Lot Number 23C100374

Method Amperometric testing with standards

| Time of Testing | Determination | Specification | Results |
|-----------------|---------------------------------|------------------|--------------|
| Initial | Average Calibration Sensitivity | 6.0-30.0 nA | 17.4 nA |
| | Average Linearity | 14.3-15.7mmol/L | 15.05 mmol/L |
| | Potassium Ferrocyanide | -0.3 - 0.3mmol/L | 0.00 mmol/L |

| \sim $^{\prime}$ | | | |
|--------------------|----|------------|----------|
| ('Antarme | tΛ | acceptance | Critaria |
| COHIDINIS | w | acceptance | Unicha |

Completed by: Christy Woodford 3/21/2023

Your Name

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.

ODON CERTIFICATION OF THE PROPERTY OF THE PROP

Gregory W. Popp Quality Manager Analytics

Dregory W. Popp