Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

a xylem brand

Product Category	Bioanalytical Standard	
Product Model Number	2790	
Lot Number	23D100172	
Method	Comparative enzymatic testing	g of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
Ethanol, 2.0 g/L	(Average of production) - (average of metrology)	± 40 mg/L	3.80 mg/L
	Coefficient of variation for production standard (CV, %)	≤ 2.0%	0.46%
Ethanol, 3.2 g/L	(Average of production) - (average of metrology)	± 64 mg/L	23.15 mg/L
	Coefficient of variation for production standard (CV, %)	≤ 2.0%	0.33%

Conforms to acceptance criteria					
Completed by:	Amanda Fenner	Date:	4/12/2023		

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp Quality Manager, Analytics

Dregory W. Popp