## **Certificate of Analysis**



## **YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category	Check Solution
Product Model Number	7173
Lot Number	23G100331
Method	Comparative conductivity testing of Production standard vs. Metrolog

<b>Chemistry Tested</b>	Determination	Tolerance	Result
Sodium	Sodium (Average of production) - (average of metrology)		0.11%
	Coefficient of variation for production standard (CV, %)	≤ 2.0%	0.36%

Conforms to acceptance criteria					
Completed by:	Amanda Fenner		Date:	7/27/2023	

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp Quality Manager, Analytics

Dregory W. Popp