Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

| Product Category | Bioanalytical Standard | |
|----------------------|-------------------------------|---|
| Product Model Number | 2790 | _ |
| Lot Number | 23H100011 | _ |
| Method | Comparative enzymatic testing | g of Production standard vs. Metrology standard |

| Chemistry Tested | Determination | Tolerance | Result |
|------------------|--|-------------------|------------|
| Ethanol, 2.0 g/L | (Average of production) - (average of metrology) | $\pm40~{ m mg/L}$ | 2.45 mg/L |
| | Coefficient of variation for production standard (CV, %) | ≤ 2.0% | 0.85% |
| Ethanol, 3.2 g/L | (Average of production) - (average of metrology) | ± 64 mg/L | 13.30 mg/L |
| | Coefficient of variation for production standard (CV, %) | ≤ 2.0% | 0.73 % |

| Conforms to acceptance criteria | | | | | |
|---------------------------------|---------------|-------|-----------|--|--|
| Completed by: | Amanda Fenner | Date: | 8/04/2023 | | |

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp Quality Manager, Analytics