Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category Membrane

Product Model Number 2329

Lot Number 23H100367

Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	6.0-30.0 nA	15.6 nA
	Average Linearity	14.3-15.7mmol/L	14.77 mmol/L
	Potassium Ferrocyanide	-0.3 - 0.3mmol/L	0.01 mmol/L

Conforme	to	acceptance	critoria
COHIDINIS	ιυ	acceptance	Cillella

Completed by: Christy Woodford 8/22/2023

Your Name

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.

OODI CERTIFICATION OF THE PROPERTY OF THE PROP

Gregory W. Popp Quality Manager Analytics

Dregory W. Popp