Certificate of Analysis



YSI Incorporated Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category	Buffer and solid products		
Product Model Number	1579		

Lot Number 23J100098

Method Verification of mass and pH (for buffers)

Determination	Specification	Result
Average mass	N/A	N/A
Average pH (buffers only)	9.00 ± 0.10	8.97

Conforms to accepta	nce criteria			
Completed by:	Amanda Fenner	Da	ate:	9/06/2023

 $\label{thm:conduct} \textit{The procedures used to manufacture and test this product have been verified by \textit{QA} for \textit{GMP compliance}. \\$



Gregory W. Popp Quality Manager, Analytics

Dregory W. Popp