Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

| Product Category | Bioanalytical Standard | |
|----------------------|------------------------|--|
| Product Model Number | 2790 | |

Lot Number 23J100241

Method Comparative enzymatic testing of Production standard vs. Metrology standard

| Chemistry Tested | Determination | Tolerance | Result |
|------------------|--|-------------------|------------|
| Ethanol, 2.0 g/L | (Average of production) - (average of metrology) | $\pm40~{ m mg/L}$ | 8.85 mg/L |
| | Coefficient of variation for production standard (CV, %) | ≤ 2.0% | 0.88 % |
| Ethanol, 3.2 g/L | (Average of production) - (average of metrology) | ± 64 mg/L | 19.65 mg/L |
| | Coefficient of variation for production standard (CV, %) | ≤ 2.0% | 0.92 % |

| Conforms to accep | otance criteria | | | |
|-------------------|-----------------|-------|-----------|--|
| Completed by: | Amanda Fennerj | Date: | 9/13/2023 | |

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp Quality Manager, Analytics