Certificate of Analysis



Product Category

YSI Incorporated Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Membrane

Product Model Number

2761

Lot Number

23M100365

Method

Amperometric testing with standards

Time of Testing	Determination	Specification	Result
Initial	Average calibration sensitivity	10.0 - 60.0 nA	34.21 nA
	Average Linearity	27.0 – 33.0 g/L	29.88 g/L
	Potassium Ferrocyanide interference	0 - 0.07 g/L	0.00 g/L
Day =	Average calibration sensitivity	N/A	N/A
	Average Linearity	N/A	N/A
	Potassium Ferrocyanide interference	N/A	N/A

The procedures used to manufacture and test this product have been verified by QA for compliance. The electronic signature below is representative of the Chemistry Technician testing and verifying conformity to acceptance criteria.

Completed by: Kristopher Kleski Date: 12/22/2023

