



# ***Certificate of Analysis***

## **Conductivity Calibration Solution**

### **3163**

1725 Brannum Lane  
Yellow Springs, Ohio  
45387-1107 USA  
+1 937-767-7241

#### Calibration Equipment:

- IET 1693 RLC Digibridge  
Instrument #: 7120  
Calibration Due Date: 4/30/2026
- Agilent 34401A 6½ Digit Multimeter  
Instrument #: 6859  
Calibration Due Date: 10/31/2025

## **Testing Results**

Lot Number: 25E100249

Nominal Conductivity  
(microsiemens/cm):

10K

Specification Range  
(microsiemens/cm):

9975 - 10025

Test Result  
(microsiemens/cm):

9993

#### Traceability:

We certify that we have calibrated this product following manufacturing standards with calibrations that are traceable to the National Institute of Standards and Technology (NIST) or are derived from accepted values of natural physical constants. As manufactured, the product meets or exceeds its current specifications.

Technician

Date

Kaity Sisco

5/12/2025