



# ***Certificate of Analysis***

1725 Brannum Lane  
Yellow Springs, Ohio  
45387-1107 USA  
+1 937-767-7241

Product Category	ISE Reagent
Product Model Number	2971 (1000 mg/L Potassium, 10 mg/L Ammonium)
Lot Number	25F100051
Method	Comparative ISE testing of Production reagent vs. Metrology reagent

## **Testing Results**

Chemistry Tested	Determination	Tolerance	Result
Potassium	(Average of Production) – (Average of Metrology)	$\leq 20$ mg/L (2%)	2.00
	Coefficient of Variation (CV) for Production reagent	$\leq 3\%$	0.78

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.

Technician

Amanda Fenner

Date

6/5/2025